

C E R T I F I C A T E
of Conformity



Registration No.: AK 50291799 0001

Report No.: 17040457 001

Holder: Caretium Medical Instruments
Co., Limited
Beishan Industrial Park 7th Floor
Building 1, Beishan Road, Yantian
Shenzhen, Guangdong 518083
P.R. China

Product: Analysis Equipment
Semi-Auto Chemistry Analyzer

Identification: Type Designation: NB-201
Serial No. : n.a.

Remark: Refer to test report 17040457 001 for details.

Tested acc. to: EN 61010-1:2010
IEC 61010-1:2010
EN 61010-2-081:2002+A1
IEC 61010-2-081:2001+A1
EN 61010-2-101:2002
IEC 61010-2-101:2002

The certificate of conformity refers to the above mentioned product. This is to certify that the specimen is in conformity with the assessment requirement mentioned above. This certificate does not imply assessment of the production of the product and does not permit the use of a TÜV Rheinland mark of conformity.

Date 11.09.2014

Certification Body
A circular purple stamp with the TÜV Rheinland logo in the center. The text around the logo reads "TÜV Rheinland LGA Products GmbH" and "Zertifizierungsstelle".
Dipl.-Ing. (FH) C. Padel

TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg