

ATTESTATION OF CONFORMITY

Attestation Number:SZ4220309-07992E-SF-01

Date of Issue:2022-06-17

Client:

Company name: Sichuan Shuke Instrument Co.,Ltd
Address: Floor 1-2, Building 12, No. 1919, Shuangyan Road, Chengdu
Cross Strait Science and Technology Industrial Development Park,Wenjiang
District,Chengdu City,Sichuan,China.

Product:

Name: Refrigerated Centrifuge
Model(s): LG-22, LG-18, LD-5M, LD-6M, LD-6, LD-5G, L5KR, TGL-25, TGL-23, TGL-22,
TGL-21, TGL-22S, TGL-20, TGL-19, TGL-17, TGL-18, TGL-1650, TGL-16,
TGL-16S, TGL-2150, TGL-1850, TDL-6M, TDL-5M, TDL-6, TDL-550, TDL-5,
TDL-4)

Manufacturer & Address: Sichuan Shuke Instrument Co.,Ltd&Floor 1-2, Building 12, No. 1919,
Shuangyan Road, Chengdu Cross Strait Science and Technology Industrial
Development Park,Wenjiang District,Chengdu City,Sichuan,China.

Trade Mark:

N/A

Bay Area Compliance Laboratories Corp. (Shenzhen) hereby declares that the submitted sample(s) of the above equipment has been tested for CE regulations and in accordance with the European Directives and Standards:

Regulation (EU) 2017/746 on in vitro diagnostic medical medical devices

Harmonized Standards	Test Report Number
EN 61010-1:2010+A1:2019 EN IEC 61010-2-010:2020 EN 61010 2-020:2017 EN61010-2-101:2017	SZ4220309-07992E-SF-01
EN 13532:2002	SZ4220309-07992E-SF-02
EN IEC 61326-1:2021 EN IEC 61326-2-6:2021	RSZ220323003-1



Mark is permitted only after all applicable requirements are met in accordance with the CE regulation requirements, including the manufacturer’s issuance of a “Declaration of Conformity. The Declaration of Conformity is issued under the sole responsibility of the manufacturer. This attestation is specific to the standard(s) stated above and compliance with additional standards and/or CE regulations are applicable.

Attestation by:

Lab. Manager: Jeanne Han

Signature:

